



SGS

EU Type Examination Certificate Number: **0120/SGS0435**

Zhejiang CHINT Instrument & Meter Co., Ltd

Wenzhou Bridge Industrial Zone
Yueqing
Zhejiang
P.R. China

Instrument Identification:
DTSU666, DTSU666-CT

**Polyphase, Active Import/Export kWh, DIN Rail, Transformer Operated/Direct Connected,
Smart, Electricity Meter**

Instrument Traceable Number
0120/SGS0435

has been assessed and certified as meeting the requirements of

EU Directive 2014/32/EU

on Measuring Instruments Annex II, Module B

It is certified that the manufacturer's technical design and specimen for the above instrument has been examined and, based on the evidence submitted, it is considered that the instrument conforms to the requirements of Annex V of EU Directive 2014/32/EU

This certificate must be used in conjunction with a certificate covering the product verification as required in Annex II, Module D or Annex II, Module F

This certificate is valid for 10 years from 23rd January 2020 until 23rd January 2030
Issue 1

Certification is based on report number(s)
SHES190501593701 dated 13th January 2020
EMA273238/1

Authorised Signature

SGS United Kingdom Limited, Notified Body 0120
Unit 202B Worle Parkway, Weston-super-Mare, BS22 6WA, UK
t +44 (0)1934 522917 f +44 (0)1934 522137 www.sgs.com

Contact Address
SGS United Kingdom Limited, Units 12A & 12B, South Industrial Estate, Bowburn, Durham, DH6 5AD, UK
t +44 (0)191 377 2000 f +44 (0)191 377 2020 www.sgs.com




This document is issued, on the Client's behalf, by the Company under its General Conditions of Service printed overleaf. The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

Any other holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Clients instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents.

EU Type Examination Cert.


SGSPAPER
19794739




	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020

1. Technical Data


Manufacturer	Zhejiang CHINT Instrument & Meter Co., Ltd.
Meter Type	DTSU666 DTSU666-CT
Voltage Rating (<i>Un</i>)	3*220/380V 3*230/400V 3*240/415V
Current Rating (<i>I_{min}</i> – <i>I_{ref}</i> (<i>I_{max}</i>))	DTSU666: 0.25-5(80)A DTSU666-CT: 0.015-1.5(6)A
Frequency (<i>Fn</i>)	50Hz
Active Accuracy Class (<i>kWh</i>)	DTSU666: Class B DTSU666-CT: Class C
Type of circuit	3P4W, 3P3W, 1P2W
Temperature Range	-25°C to +55°C
Software/ Firmware Version No's	DTSU666: DTSU666 5(80)A V108 DTSU666-CT: DTSU666-CT 1.5(6)A V128
CRC Checksum No's	DTSU666: 0108 DTSU666-CT: 0128
Identification Location	LCD
Bill Of Materials No's	DTSU666: BOM-DTSU666 5(80)A V2.1 DTSU666-CT: BOM-DTSU666-CT 1.5(6)A V2.1
IP Rating	IP51
Insulation Protective Class	Class II
LED Pulse Constant	DTSU666: 400imp/kWh DTSU666-CT: 6400imp/kWh
Impulse Voltage Rating	6kV
AC Voltage Rating	4kV
Main Cover Sealing Type	Tamper proof sealing tape
Integrity of meter	Inaccessible without breaking seals
Intended Location of the Meter	Indoor
Type of Register	LCD
Terminal Arrangement(s)	DIN
Location of Manufacturers Address	Associated Documents

	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020


2. Photograph of Meter and Sealing Plan



Terminal cover sealing points





Main cover sealing point

	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020

3. Examples of Nameplates

CHNT

DTSU666 Three-Phase Smart Meter






50/60Hz

400imp/kWh

0.25-5(80)A



3X220/380V...3X240/415V EN50470-1\3 Cl.B 3K6

SET



ESC

→

0120/SGS0435   M19 0120

CHNT

DTSU666-CT Three-Phase Smart Meter






50/60Hz

6400imp/kWh

0.015-1.5(6)A



3X220/380V...3X240/415V EN50470-1\3 Cl.C 3K6





SET

ESC

→

0120/SGS0435   M19 0120

	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020


4. Calculation of the composite error/ MPE

During the type approval examination the influence factors for temperature, frequency and voltage are determined per load point. The table below represents the sum of the square values per load, determined via the following formula:-

$$\delta e(T, U, f) = \sqrt{(\delta e^2(T, I, \cos\phi) + \delta e^2(U, I, \cos\phi) + \delta e^2(f, I, \cos\phi))}$$

where

- $\delta e(T, I, \cos\phi)$ = Additional error due to variation of the temperature at the same load
- $\delta e(U, I, \cos\phi)$ = Additional error due to variation of the voltage at the same load
- $\delta e(f, I, \cos\phi)$ = Additional error due to variation of the frequency at the same load

	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020

DTSU666

		Influence Factors for Temperature, Frequency & Voltage					
Current	PF Cos	-25°C	-10°C	5°C	30°C	40°C	55°C
I _{min}	1.0	0.13	0.13	0.09	0.05	0.11	0.22
I _{tr}	1.0	0.15	0.15	0.12	0.02	0.10	0.23
10I _{tr}	1.0	0.16	0.16	0.12	0.03	0.11	0.26
I _{max}	1.0	0.21	0.21	0.17	0.03	0.06	0.22
I _{tr}	0.5ind	0.14	0.15	0.13	0.04	0.08	0.20
10I _{tr}	0.5ind	0.12	0.13	0.10	0.04	0.10	0.25
I _{max}	0.5ind	0.25	0.25	0.22	0.09	0.10	0.31
I _{tr}	0.8cap	0.17	0.16	0.13	0.04	0.09	0.20
10I _{tr}	0.8cap	0.18	0.17	0.14	0.05	0.12	0.27
I _{max}	0.8cap	0.22	0.21	0.18	0.06	0.05	0.18
L1							
I _{tr}	1.0	0.19	0.19	0.16	0.02	0.08	0.22
10I _{tr}	1.0	0.18	0.16	0.13	0.04	0.12	0.28
I _{max}	1.0	0.21	0.20	0.17	0.02	0.08	0.24
I _{tr}	0.5ind	0.17	0.18	0.16	0.04	0.07	0.22
10I _{tr}	0.5ind	0.12	0.12	0.10	0.04	0.11	0.28
I _{max}	0.5ind	0.26	0.24	0.19	0.04	0.09	0.35
L2							
I _{tr}	1.0	0.09	0.10	0.09	0.03	0.09	0.23
10I _{tr}	1.0	0.12	0.13	0.11	0.01	0.08	0.21
I _{max}	1.0	0.17	0.19	0.16	0.04	0.03	0.17
I _{tr}	0.5ind	0.05	0.06	0.06	0.05	0.12	0.25
10I _{tr}	0.5ind	0.10	0.10	0.10	0.03	0.07	0.18
I _{max}	0.5ind	0.26	0.26	0.23	0.15	0.13	0.23
L3							
I _{tr}	1.0	0.17	0.16	0.12	0.04	0.11	0.24
10I _{tr}	1.0	0.18	0.17	0.12	0.05	0.13	0.30
I _{max}	1.0	0.21	0.20	0.15	0.03	0.11	0.28
I _{tr}	0.5ind	0.11	0.10	0.07	0.06	0.15	0.22
10I _{tr}	0.5ind	0.16	0.15	0.12	0.03	0.15	0.28
I _{max}	0.5ind	0.24	0.27	0.11	0.05	0.17	0.29



EU-Type Examination Certificate Number:


0120/SGS0435

Issue Number: 1

Dated: 23rd January 2020

DTSU666-CT

		Influence Factors for Temperature. Frequency & Voltage					
Current	PF Cos	-25°C	-10°C	5°C	30°C	40°C	55°C
I _{min}	1.0	0.02	0.04	0.04	0.03	0.08	0.21
I _{tr}	1.0	0.01	0.05	0.06	0.02	0.08	0.21
10I _{tr}	1.0	0.02	0.05	0.06	0.02	0.08	0.21
I _{max}	1.0	0.01	0.06	0.06	0.02	0.08	0.21
I _{tr}	0.5ind	0.05	0.03	0.05	0.02	0.09	0.26
10I _{tr}	0.5ind	0.10	0.09	0.08	0.08	0.12	0.26
I _{max}	0.5ind	0.06	0.37	0.05	0.05	0.16	0.23
I _{tr}	0.8cap	0.06	0.07	0.07	0.02	0.07	0.19
10I _{tr}	0.8cap	0.04	0.10	0.08	0.03	0.09	0.20
I _{max}	0.8cap	0.03	0.13	0.07	0.02	0.06	0.22
					0.00	0.00	0.00
L1							
I _{tr}	1.0	0.04	0.08	0.08	0.04	0.07	0.20
10I _{tr}	1.0	0.01	0.06	0.06	0.01	0.08	0.21
I _{max}	1.0	0.01	0.06	0.06	0.02	0.08	0.21
I _{tr}	0.5ind	0.06	0.11	0.12	0.06	0.09	0.29
10I _{tr}	0.5ind	0.10	0.09	0.10	0.08	0.12	0.29
I _{max}	0.5ind	0.01	0.07	0.07	0.01	0.07	0.25
L2							
I _{tr}	1.0	0.02	0.05	0.05	0.02	0.08	0.21
10I _{tr}	1.0	0.02	0.06	0.06	0.01	0.07	0.19
I _{max}	1.0	0.06	0.10	0.11	0.06	0.07	0.18
I _{tr}	0.5ind	0.04	0.07	0.08	0.04	0.10	0.21
10I _{tr}	0.5ind	0.22	0.06	0.10	0.06	0.08	0.19
I _{max}	0.5ind	0.40	0.38	0.46	0.38	0.34	0.35
L3							
I _{tr}	1.0	0.07	0.11	0.11	0.06	0.08	0.20
10I _{tr}	1.0	0.01	0.07	0.07	0.02	0.09	0.23
I _{max}	1.0	0.01	0.08	0.07	0.02	0.09	0.23
I _{tr}	0.5ind	0.11	0.12	0.14	0.11	0.11	0.20
10I _{tr}	0.5ind	0.09	0.04	0.05	0.05	0.07	0.22
I _{max}	0.5ind	0.04	0.05	0.05	0.01	0.06	0.19


	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020

5. Annex of Variants

Product Variant Identification Details:

Type Designation	Description of meter
DTSU666	Direct connected, 3x220/380V – 240/415V, 5(80)A
DTSU666-CT	Transformer operated, 3x220/380V – 240/415V, 1.5(6)A

Modifications to the meter(s) described according to approval No.**0120/SGS0435** must be notified to the issuing body to confirm the meter(s) continuing compliance to the relevant pattern approval standard(s).

	EU-Type Examination Certificate Number:	
	0120/SGS0435	
	Issue Number: 1	Dated: 23 rd January 2020

6. Document Revision History

Issue	Date	Comments
1	23/01/2020	Initial Issue

This document is issued by the Company subject to its General Conditions for Certification Services, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>.

Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested *and such sample(s) are retained for 28 days only.*

END OF CERTIFICATE